

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.  
Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
Person responsible: Ing. C. Oosterman

### Applicant

Name: Mettler-Toledo (Changzhou) Measurement Technology Ltd.  
Address: 111, West HaiHu Rd, XinBei District,  
ChangZhou, XinBei District  
213125 Jiangsu  
Peoples Republic of China

### Manufacturer of the certified type

Name: Mettler-Toledo (Changzhou) Measurement Technology Ltd.  
Address: 111, West HaiHu Rd, XinBei District,  
ChangZhou, XinBei District  
213125 Jiangsu  
Peoples Republic of China

### Identification of certified type

Indicator, as a part of a non-automatic weighing instrument.  
Type: IND131 / IND331

The maximum number of verification scale intervals will be:  
 $n \leq 6000$  for class **III** instruments or;  
 $n \leq 1000$  for class **IIII** instruments.

**OIML Member state**  
The Netherlands

**OIML Certificate N° R76/2006-NL1-09.26**  
Project number 9200104  
Page 2 of 2

Applied error fraction  $p_i$ : 0.5  
Minimum input impedance: 87  $\Omega$   
Maximum input impedance: 350  $\Omega$   
Signal voltage per verification scale interval  $\geq 0.6 \mu\text{V}$   
Temperature range  $-10^\circ\text{C} / +40^\circ\text{C}$

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the test certificate and the description with number TC7600 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**  
Edition 2006 (E)  
for accuracy class (III) or (IIII)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:  
N° R76/2006-NL1-09.26A that includes 38 pages;  
N° R76/2006-NL1-09.26B that includes 30 pages.

The Issuing Authority NL1  
NMI Certin, 20 August 2009



C. Oosterman  
Head Certification Board

\*  
\* \*

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.